



cellasys GmbH
Illerstrasse 14
87758 Kronburg / Germany

www.cellasys.com
info@cellasys.com

BioChip Dummy (Order No. 08815)

Description

BioChip Dummy to verify proper function of IMOLA-IVD set up.



In general **BioChip Dummy** may only be used in combination with cellasys IMOLA-IVD by **qualified personnel** of a research or healthcare institution.
Read the **IMOLA-IVD manual** thoroughly and carefully follow the instructions and guidelines provided!

BioChip Dummy values:

	Value
pH1	300 mV +/- 30 mV
pH2	300 mV +/- 30 mV
O ₂	1850 mV +/- 185 mV
Temp	1500 mV +/- 150 mV
IDES1 real	130 Ω +/- 13 Ω
IDES1 imaginary	-35 Ω +/- 5 Ω
IDES2 real	130 Ω +/- 13 Ω
IDES2 imaginary	-35 Ω +/- 5 Ω

Data sheet

Values without BioChip Dummy (open circuit):

	Value	
pH1	0 mV	+/- 30 mV
pH2	0 mV	+/- 30 mV
Oxygen	2075 mV	+/- 185 mV
Temperature	0 mV	+/- 150 mV
Real1	0 Ω	+/- 5 Ω
Imaginary1	0 Ω	+/- 5 Ω
Real2	0 Ω	+/- 5 Ω
Imaginary2	0 Ω	+/- 5 Ω

Technical data

Dimensions: 24,0 x 24,0 x 10,0 mm³

Weight: 2,6 g

Operating temperature: 0 °C to +80°C

Liability / Copyright

All technical details are state of the technology from November 2019 and are subject to change without notice. No liability is assumed for pictures, descriptions or any content of this document.

All descriptions, pictures, technical drawings and all other illustrations are protected by copyright and unless otherwise marked property of cellasys GmbH.

Any subsequent use needs prior written, allowance by cellasys GmbH.